

Frank Hertz Experiment

Introduction


From the early spectroscopic work it is clear that atoms emitted radiation at discrete frequencies; from Bohr's model, the frequency of the radiation ν is related to the change of energy levels through $DE=h\nu$. It is then to be expected that transfer of energy to atomic electrons by any mechanism should always be in discrete amounts. One such mechanism of energy transfer is through inelastic scattering of low-energy electrons.



Frank and Hertz in 1914 set out to verify these considerations.


- It is possible to excite atoms by low energy electron bombardment.
- The energy transferred from electrons to the atoms always had discrete values.
- The values so obtained for the energy levels were in agreement with spectroscopic results.

Thus the existence of atomic energy levels put forward by Bohr can be proved directly. It is a very important experiment and can be performed in any college or University level laboratory.



Operating Principle

The Frank-Hertz tube in this instrument is a tetrode filled with the vapour of the experimental substance. Fig.1 indicates the basic scheme of experiment. The electrons emitted by filament can be accelerated by the potential V_{G2K} between the cathode and the grid G_2 . The grid G_2 helps in minimizing space charge effects. The grids are wire mesh and allow the electrons to pass through. The plate A is maintained at a potential slightly negative with respect to the grid G_2 .



This helps in making the dips in the plate current more prominent. In this experiment, the electron current is measured as a function of the voltage V_{G2K} . As the voltage increases, the electron energy goes up and so the electron can overcome the retarding potential V_{G2A} to reach the plate A . This gives rise to a current in the ammeter, which initially increases. As the voltage further increases, the electron energy reaches the threshold value to excite the atom in its first allowed excited state. In doing so, the electrons lose energy and therefore the number of electrons reaching the plate decreases. This decrease is proportional to the number of inelastic collisions that have occurred. When the V_{G2K} is increased further and reaches a value twice that of the first excitation potential, it is possible for an electron to excite an atom halfway between the grids, lose all its energy, and then gain enough energy to excite a second dip in the current. The advantage of this type of configuration of the potential is that the current dips are much more pronounced, and it is easy to obtain fivefold or even larger multiplicity in the excitation of the first level.

Frank-Hertz Experiment Set-up, Model: FH-3001, consists of the following:

- Argon filled tetrode
- Filament Power Supply : 2.6-3.4V continuously variable
- Power Supply for VG1K : 1.3-5V continuously variable
- Power Supply for VG2A : 1.3 - 12V continuously variable
- Power Supply for VG2K : 0 - 95V continuously variable

All the power supplies are highly stabilized and output voltages can be read on 3 1/2 digit, 7 segment LED DPM with auto polarity and decimal indication through a selector switch.





- Saw tooth waveform for CRO display
- Scanning Voltage : 0-80V
- Scanning Frequency : 115±20Hz
- Multirange Analogue Voltmeter
- Range : 0-5V, 0-15V & 0-100V
- Multirange Digital Ammeter
- Display : 3 1/2 digit 7 segment LED
- Range Multiplier : 10⁻⁷, 10⁻⁸, 10⁻⁹
- Power : 220V±10% mains, 50Hz.

The instrument can, not only lead to a plot of the amplitude spectrum curve by means of point by point measurement, but also directly display the amplitude spectrum curve on the oscilloscope screen. This instrument can thus be used as a classroom experiment as well as for demonstration to a group of students.

Analysis of Data

Data obtained for the excitation potential point by point are shown in Fig. 3. The readings are taken for 1V changes on grid 2 (VG2K). A significant decrease in electron (collector) current is noticed every time the potential on grid 2 is increased by approximately 12V, thereby indicating that energy is transferred from the beam in (bundles) "quanta" of 12eV only. Indeed, a prominent line in the spectrum of argon exists at 1048Å corresponding to eV=11.83.

The location of the peaks is indicated in Fig. 3. Average value of spacing between peaks is 11.75eV compared with the accepted value of 11.8

**CROMTECH INDIA**

1st Floor Begum Niwas, Opp Mercantile, Bank Society Colony, 5th Natwar Nagar Road, Jogeshwari (East), Mumbai 400 060.
Maharashtra - India.

Mobile: +91 98693 82134 / 81088 13320

Telefax: +91 22 2838 6028

Email: cromamumbai@mail.com / cromtechindia@mail.com / info@cromtechindia.com / sales@cromtechindia.com

Website: www.cromtechindia.com